User Guide:
Nebulising with the PARI LC SPRINT®

Please ensure you read the Instructions For Use that come with the nebuliser and compressor, which give a more detailed description than this short guide. The instruction manual is included in the device package. If lost you can download the manual at www.pari.de/en or request it by calling 01932 341122.

1

Always wash hands thoroughly before use.

2a

Attach the mouthpiece or mask.

2b

If using a BABY mask, always use the BABY bend and position the mask according to the position of the baby.

3

Connect the nebuliser and compressor with the connection tubing. Connect the compressor to the mains power supply.

4

Open the cap.

5

Fill the nebuliser with the amount of medication prescribed by your doctor.

6

Sit up and relax.
Switch on device.

7a

Place the mouthpiece between the teeth and close the lips around it, or place the mask tightly over the mouth and nose.

7b

Always keep the nebuliser vertical. For babies gently press the mask over the mouth and nose.

8

Breathe in slowly and deeply through the mouthpiece or mask. Breathe out again slowly and fully. When a different noise comes from the nebuliser, this indicates nebulisation has finished.

9

Switch off device. Disconnect the nebuliser from the tubing. Switch the device back on to dry the tubing.

10



Disconnect the power plug.

User Guide:
Nebulising with the PARI LC SPRINT®

Please ensure you read the Instructions For Use that come with the nebuliser and compressor, which give a more detailed description than this short guide. The instruction manual is included in the device package. If lost you can download the manual at www.pari.de/en or request it by calling 01932 341122.

Advice on hygiene when nebulising at home

1

Cleaning

- Dismantle the nebuliser into separate parts.
- Wash all parts with warm water and some washing up liquid.
- Thoroughly rinse all parts under running water.

2

Disinfecting

- In a sufficient amount of boiling water (for at least 5 minutes).
- Vaporisator (steam disinfector).

3

Drying

- Place all parts on a dry, lint-free cloth (for at least 4 hours). Before re-assembling, all parts must be completely dry.

4

Storage

- Store parts in a dry, dust-free environment.

5

Maintenance: at least once a year

- Change nebuliser, air filter and connection tubing (PARI Year Pack). If dirty or damaged, replace sooner.

© PARI GmbH · Spezialisten für effektive Inhalation · 023D0042-C-05/2012

PARI Medical Ltd
The Old Sorting Office, Rosemount Avenue,
West Byfleet, Surrey KT14 6LB
Tel: 01932 341122
Fax: 01932 341134
E-Mail: infouk@pari.eu
www.parimedical.co.uk

Specialists in effective inhalation